

TECNIS Synergy[™] IOL

INDICATIONS AND IMPORTANT SAFETY INFORMATION FOR TECNIS SYNERGYTM OPTIBLUE AND TECNIS SYNERGY™ TORIC II OPTIBLUE™ IOL WITH TECNIS SIMPLICITY™ DELIVERY SYSTEM, MODEL DFW

INDICATIONS FOR USE

The TECNIS Simplicity™ Delivery System is used to fold and assist in inserting the TECNIS Synergy™ OptiBlue™ IOL which is indicated for primary implantation for the visual correction of aphakia in adult patients with or without presbyopia, in whom a cataractous lens has been removed. The lens mitigates the effects of presbyopia by providing vision far through near and reduced spectacle dependence across a range of distances

The TECNIS Simplicity™ Delivery System is used to fold and assist in inserting the TECNIS Synergy™ Toric II OptiBlue™ IOLs which are indicated for primary implantation for the visual correction of aphakia and pre-existing corneal astigmatism in adult patients with or without presbyopia, in whom a cataractous lens has been removed. The lens mitigates the effects of presbyopia by providing vision far through near, a reduction of residual refractive cylinder and reduced spectacle dependence across a range of distances. The lenses are intended to be placed in the capsular bag.

CONTRAINDICATIONS None

RISKS

Routine cataract surgery risk, irrelevant to lens selection, could be minor, temporary, or affect patients' vision permanently. Rare complications are worsening of vision, bleeding, or infection. Risks related to use of this lens include halo or glare in nighttime or poor visibility conditions. These may be perceived as an annoyance or hindrance, which on rare occasions, may be significant enough for the patient to request removal of the IOL. Patients should discuss all risks and benefits with their eye doctor before surgery.

PRECAUTIONS

Recent contact lens usage may affect the patient's refraction; therefore, in contact lens wearers, surgeons should establish corneal stability without contact lenses prior to determining IOL power. Prior to surgery, the surgeon must inform prospective patients of the possible risks and benefits associated with the use of this device and provide a copy of the patient information brochure to the patient. Safety and effectiveness in patients 21 years or younger have not been established in clinical studies. A patient implanted with an intraocular lens should be monitored on a regular basis for long-term postoperative follow-up. The intraocular pressure of implanted patients with glaucoma should be carefully monitored for postoperative changes.

WARNINGS

Physicians considering lens implantation should weigh the potential risk/benefit ratio. Well-informed patients with well-defined visual needs and preferences should be selected for lens implantation. The patients should be informed of the possibility of visual effects (such as halo or glare) in nighttime or poor visibility conditions. Patients may perceive these visual effects as an annoyance or hindrance, which, on rare occasions, may be significant enough for the patient to request removal of the IOL. Patients with a predicted postoperative astigmatism greater than 1.0 diopter may not be suitable candidates for lens model ZFW implantation since they may not fully benefit in terms of potential spectacle independence. The lens may affect image quality and lead to some reduction of contrast sensitivity compared to a monofocal lens. Therefore, patients should exercise caution when driving at night or in poor visibility conditions.

ATTENTION

Reference the Directions For Use for a complete listing of indications and safety information.

PLEASE CONSULT WITH YOUR DOCTOR ABOUT COMPLICATIONS AND RISKS BEFORE SURGERY.

- 1. TECNIS Synergy IOL DFU Canada Doc# Z311435 REF2020CT4195
- 2. DOF2019OTH4003_TECNIS Synergy 6-Month POC Data
- 3. DOF2019OTH4002 TECNIS Synergy IOL, model ZFR00V: MTF bench data compared with other lenses
- 4. Presbyopia. American Optometric Association Website: https://www.aoa.org/patients-andpublic/eye-and-vision-problems/glossary-of-eye-and-vision-conditions/presbyopia
- 5. TECNIS Synergy™ Toric II OptiBlue™ DFU-Canada-Doc# Z311562E Rev. A 04/2021. REF2021CT4147
- 6. TECNIS Synergy™ OptiBlue™ IOL with TECNIS Simplicity™ DFU-Canada-Doc# Z311561E Rev. B 03/2021, REF2021CT4106

Johnson Johnson vision

©Johnson & Johnson Vision, Inc. 2021 PP2021CT4454

FROM FAR THROUGH NEAR, DAY AND NIGHT.



The Continuous-Range-of-Vision¹ IOL from Johnson & Johnson Vision

LIVE THE LIFE **YOU WANT TO SEE.**

TECNIS Synergy IOL



Johnson Johnson vision

With **TECNIS Synergy™** IOLs, you have the opportunity to see a continuous range of high-quality vision, even up

close and in low light.^{2,3}



Presbyopia is a gradual, natural blurring of sight at close distances. You might already be aware of presbyopia as the reason you need to wear reading glasses.

But here's something you may not know: cataract surgery can be a great opportunity to address your presbyopia along with your cataract, with or without astigmatism.

YOUR LIFE ISN'T LIVED AT ONE DISTANCE.

A range of vision that's as dynamic as your life.

From working at computer screens to reading in low-light, our vision needs change from moment to moment. What if you could see clearly from all distances and levels of light?





With presbyopia



With **TECNIS Synergy™** IOL

These images are for illustrative purposes only and do not represent actual data derived from studies. These illustrative simulations are intended to help you better understand your vision in certain eye conditions.

PRESBYOPIA



Typically becomes noticeable around the age of 40^4



Makes it difficult to focus on things up close (near vision)



Can be addressed with a presbyopiacorrecting IOL

Other solutions include reading glasses, contacts or an additional surgery

BENEFITS OF TECNIS SYNERGY™ IOLS



Provides a continuous range of vision that allows you to focus far through near.²



Delivers high-quality vision for clarity day and night.³

Talk to your doctor about the life you want to see, and find out if **TECNIS Synergy**[™] IOLs are right for you.

Johnson Johnson vision